

HUMAN

Target Values

BioRad Lyphocheck Immunoassay Plus Control

REF 370, **LOT** 40330 Level 1 - 3  2019-05-31

REF		Units Einheiten Unitades Unités	Target Sollwert Valor meta Valeur souhaitée	Range Bereich Rango Marge	Target Sollwert Valor meta Valeur souhaitée	Range Bereich Rango Marge	Target Sollwert Valor meta Valeur souhaitée	Range Bereich Rango Marge
HUMAN Hormones & Tumor Markers								
			Level 1 - 40331		Level 2 - 40332		Level 3 - 40333	
52010	AFP	ng/ml	18.6	12.5 - 24.8	86.49	58.0 - 115	202	135 - 268
52020	CEA	ng/ml	1.91	1.28 - 2.54	14.36	9.62 - 19.1	21.4	14.3 - 28.4
55050	Cortisol	ng/ml	41.7	26.7 - 56.7	246	157 - 335	390.8	250 - 532
55060	DHEA-S	µg/ml	0.51	0.28 - 0.74	1.40	0.77 - 2.03	4.82	2.65 - 6.99
55030	Estradiol	pg/ml	162	90.7 - 233	449	252 - 647	872	488 - 1256
55040	Estriol	ng/ml	1.78	0.98 - 2.57	5.76	3.17 - 8.35	12.9	7.08 - 18.68
52035	fPSA	ng/ml	0.4	0.27 - 0.53	1.93	1.30 - 2.57	>9	-
53020	FSH	IU/l	7.77	5.20 - 10.33	18.3	12.2 - 24.3	33.1	22.2 - 44.0
54015	fT3	pg/ml	2.63	1.76 - 3.49	4.94	3.31 - 6.57	7.05	4.72 - 9.37
54025	fT4	ng/dl	0.79	0.53 - 1.05	2.07	1.39 - 2.75	3.86	2.59 - 5.14
53040	hCG	IU/l	6.17	4.13 - 8.20	23.9	16.0 - 31.8	205.1	137 - >250
53010	LH	IU/l	1.37	0.92 - 1.82	19.7	13.2 - 26.2	58.6	39.3 - 77.9
53030	Progesterone	ng/ml	0.97	0.53 - 1.40	11.68	6.43 - 16.9	32.9	18.1 - 47.8
55020	Prolactin	ng/ml	7.02	4.70 - 9.34	19.2	12.9 - 25.5	44.6	29.9 - 59.3
52030	PSA	ng/ml	1.17	0.90 - 1.43	5.51	4.30 - 6.72	31.0	24.5 - 37.5
54010	T3	ng/ml	0.91	0.61 - 1.22	2.01	1.35 - 2.67	3.19	2.14 - 4.24
54020	T4	µg/dl	4.95	3.32 - 6.59	10.3	6.88 - 13.7	16.96	11.4 - 22.6
55010	Testosterone	ng/ml	2.06	1.20 - 2.93	7.15	4.15 - 10.2	13.77	7.99 - 19.6
54030	TSH	mIU/l	0.32	0.16 - 0.48	4.51	2.63 - 6.39	*	-

* Control is not suitable for this parameter as the target value is out of the linearity range of the test.

