

HUMAN
Target Values
BioRad Lyphocheck Immunoassay Plus Control
[REF] 370, [LOT] 40340 Level 1 - 3 📅 **2019-11-30**

[REF]		Units Einheiten Unités	Target Sollwert Valor meta Valeur souhaitée	Range Bereich Rango Marge	Target Sollwert Valor meta Valeur souhaitée	Range Bereich Rango Marge	Target Sollwert Valor meta Valeur souhaitée	Range Bereich Rango Marge
HUMAN Hormones & Tumor Markers								
			Level 1 - 40341		Level 2 - 40342		Level 3 - 40343	
52010	AFP	ng/ml	13.38	8.97 - 17.8	79.0	52.9 - 105	186	125 - 248
52020	CEA	ng/ml	1.75	1.17 - 2.32	12.1	8.09 - 16.1	17.7	11.8 - 23.5
55050	Cortisol	ng/ml	36.5	23.4 - 49.6	202.5	130 - 275	308.5	197 - 420
55060	DHEA-S	µg/ml	0.42	0.23 - 0.61	1.19	0.66 - 1.73	4.08	2.24 - 5.92
55030	Estradiol	pg/ml	230	129 - 332	601.6	337 - 866	1093	612 - 1574
55040	Estriol	ng/ml	2.41	1.33 - 3.50	5.89	3.24 - 8.54	13.7	7.54 - 19.9
52035	fPSA	ng/ml	0.29	0.19 - 0.39	1.30	0.87 - 1.72	>10	-
53020	FSH	IU/l	4.25	2.85 - 5.65	18.7	12.5 - 24.9	38.6	25.9 - 51.4
54015	ft3	pg/ml	2.61	1.75 - 3.47	4.87	3.26 - 6.47	7.45	5.00 - 9.91
54025	ft4	ng/dl	0.94	0.63 - 1.25	2.27	1.52 - 3.02	3.81	2.55 - 5.07
53040	hCG	IU/l	6.92	4.63 - 9.20	19.0	12.7 - 25.3	176.4	118 - 235
53010	LH	IU/l	1.46	0.98 - 1.94	20.5	13.8 - 27.3	65.4	43.8 - 87.0
53030	Progesterone	ng/ml	0.94	0.52 - 1.36	11.4	6.27 - 16.5	33.2	18.3 - 48.1
55020	Prolactin	ng/ml	7.30	4.89 - 9.71	20.3	13.6 - 27.0	41.3	27.7 - 55.0
52030	PSA	ng/ml	0.91	0.50 - 1.32	3.23	2.10 - 4.36	26.5	17.2 - 35.8
54010	T3	ng/ml	1.01	0.68 - 1.34	2.01	1.35 - 2.68	3.31	2.21 - 4.40
54020	T4	µg/dl	5.58	3.74 - 7.42	12.2	8.14 - 16.2	16.86	11.3 - 22.4
55010	Testosterone	ng/ml	1.87	1.08 - 2.66	5.54	3.22 - 7.87	9.98	5.79 - 14.2
54030	TSH	mIU/l	0.33	0.18 - 0.48	4.69	3.05 - 6.34	*	-

* Control is not suitable for this parameter as the target value is out of the linearity range of the test.

